



PI Calibration Laboratory

A division of Politech Instruments

303, K. K. Gupta Industrial Estate, Dr. R. P. Road, Mulund (W), MUMBAI 400080

Tel : (022) 25641902 +91 7400094065 Fax (022) 25641905

Web: www.picallab.com email:politech.callab@gmail.com

TRUE COPY



CC-2528

DISCIPLINE : MECHANICAL

Certificate of Calibration

Certificate No: PICAL/1223/P/166	Date of Issue: 01.12.2023	ULR: CC25282300001194F
National Centre For Quality Calibration		
Customer Name & Address: 4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380058		

Instrument (UUC) Details			
Date of Receipt	01.12.2023	Condition when received	Satisfactory
Name	Digital Pressure Calibrator	Range	0 to 20 mbar
Location	NA	Accuracy	± 0.1% Full Scale
Make/Model No.	POLLTECH & PSI PC.	Resolution/L.C	0.001 mbar
Sr.No:	10113	Identification no.	NCQC/M-86

Calibration Details			
Job No.	CMR/4-L/1223/P/137	Parameter of Measurement	Pressure
Calibration Procedure No.	PICAL/CP/MP/01A	Calibration Method	By Comparison Method
Calibration Date	01.12.2023	Calibration Media	Air
Next Due Date (as per customer request)	01.12.2024	Place of Calibration	At Lab On-Site

Calibration carried out by (Calibration Engineer) Surekha T Lagad.

Environmental Conditions under which Calibrated		
Temperature: 23.52°C to 23.76°C	Humidity: 55.15% RH	Ambient Pressure: 100.56 kPa A

Reference Standard Equipments used				
Name	ID.No:	Certificate No:	Certified By	Validity
Digital Pressure Calibrator	PICAL/M/P/14	CC/PRL/0128/23-24	IDEMI CC-2287	01.08.2024

RESULTS OF CALIBRATION										
Sl.No	Set Pressure on UUC	GAUGE PRESSURE Reference Standard Readings						Average of Reference Standard Readings	Error ± % of Full Scale	Expanded Uncertainty at k=2
		Cycle 1		Cycle 2		Cycle 3				
		UP	DOWN	UP	DOWN	UP	DOWN			
	mbar	mbar	mbar	mbar	mbar	mbar	mbar	mbar	%	mbar
1	0.000	0.000	0.006	0.006	0.004	0.003	0.000	0.003	-0.02	0.004
2	2.000	2.017	2.015	2.016	2.015	2.014	2.013	2.015	-0.08	0.004
3	4.000	4.015	4.013	4.014	4.012	4.011	4.009	4.012	-0.06	0.004
4	6.000	6.012	6.010	6.009	6.010	6.011	6.008	6.010	-0.05	0.004
5	8.000	8.010	8.009	8.011	8.008	8.009	8.005	8.009	-0.04	0.004
6	10.000	10.007	10.005	10.006	10.005	10.007	10.006	10.006	-0.03	0.004
7	12.000	12.010	12.008	12.008	12.009	12.011	12.009	12.009	-0.05	0.004
8	14.000	14.008	14.009	14.006	14.005	14.007	14.009	14.007	-0.04	0.004
9	16.000	16.005	16.002	16.004	16.002	16.004	16.006	16.004	-0.02	0.004
10	18.000	18.003	18.006	18.004	18.002	18.004	18.006	18.004	-0.02	0.004
11	20.000	20.001	20.001	20.004	20.004	20.005	20.005	20.003	-0.02	0.004

Summary of Results		
The above calibrated Instrument is found to be within the limits of acc. Spec.	Conformity statement :	Not Required
While making decision rule expanded uncertainty in full taken into consideration.	Decision Rule :	Not Required

Parameter	Calibration Range	Error ± % of Full Scale	Expanded Uncertainty @ k=2 ± mbar
Pressure	0 to 20 mbar	0.08	As Above

Calibrated by

 Surekha T Lagad.
 Calibration Engineer

NCQC
 Valid up to 01-12-2024
 Reviewed

Approved by

 Dr. P.K. Arora
 Technical Manager

Note:

- All the Instruments/ Reference Standards used are traceable to national standards through reference standards and their calibrations are valid.
- Results reported are valid at the time of and under the stated conditions of the measurements. Satisfactory test/calibration in no way implies that the product so tested or equipment calibrated is approved by NABL. Result presented in this certificate relate only to the items mentioned & calibrated at PI CAL LAB.
- This certificate will not be reproduced except in full without written explanation.
- The uncertainties are for a confidence Probability not less than 95 % confidence level unless specified otherwise
- This report refers only to particular item (s) submitted for calibration.

End of Certificate
Page 1 of 1

NCQC System Certificate No. 238