



सी एस आई आर - राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY

(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद)

(Council of Scientific and Industrial Research)

राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएम एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)

डॉ. के. एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत

Dr. K. S. Krishnan Marg, New Delhi-110012, INDIA

दूरभाष/Phone : 91-11- 4560 8441, 8589, 8610, 9447, फैक्स/ Fax : 91-11- 4560 8448

ई-मेल/ E-mail: cfct@nplindia.org, वेबसाइट/ Website: www.nplindia.org



अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
FORCE PROVING INSTRUMENT

प्रमाण पत्र संख्या / Certificate No.

22021420/D1.05/C-464

दिनांक /Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for the next calibration	पृष्ठ /Page	पृष्ठों की संख्या /No. of Pages
25.02.2022	25.04.2024	1	2

1. Calibrated for:

M/s National Centre for Quality Calibration,
4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,
Iskcon-Ambli Road, Ambli, Ahmedabad – 380058.

Customer's Ref. No. & Dt.: Letter Dated Nil.

2. Description & identification of instrument:

Type: Load cell Digital indicator Sr. No.: 14626
Capacity: 3000 kN Manufacturer: Star Embedded Systems (P) I
Sr. No: 14626 Resolutions: 1 div.
Model: LED-SD1
Connector type: 4 pin Cable length = 11.63 m
Make : Star Embedded Systems (P) Ltd.

3. Environmental conditions:

Temperature: $(23 \pm 1)^\circ \text{C}$ **Relative humidity:** $(50 \pm 10) \%$

4. Standards used:

3000 kN force machine

Associated uncertainty:

$\pm 0.05\% (k=2)$

5. Traceability of standard used:

The standard used for calibration is traceable to the National Standard, which realize the units of quantities according to the International System of Units (SI).

6. Principle/Methodology of calibration and calibration procedure number NPL Calibration procedure No.: Sub-Div.#D1.05/Doc.#3/CP#FT/F-02 broadly based on IS 4169-2014.

No load output: The digital indicator was switched on for 30 minutes to warm up and stabilize for no load output before the start of the calibration. The no load output was noted (before taring) and the calibration signal was noted.

Preloading: Before the application of the calibration forces, the instrument was preloaded thrice to its maximum capacity and kept at full load for about 90 seconds.

Calibration: The sequence of the applied calibration force in compression is given below:

At 0° : Two series of calibration forces in increasing values. At 120° and 240° positions: One series of calibration forces each in increasing values. Creep test is performed by calculating the difference in output i_{30} obtained at 30s and i_{300} obtained at 300s after the removal of the maximum calibration force and express this difference as percentage of maximum deflection.

The calibration was made by using Self-aligning compression pads provided along with the instrument to ensure axial application of the force.

Between each series, the instrument was rotated along its axis so as to occupy during the calibration three position (0° , 120° & 240°) and the instrument was subjected to the full load once for about 90 seconds each time before starting in a new position.

Between the loadings, readings corresponding to no load after waiting at least 30 seconds for the return to zero were noted.

NQC
Valid up to 25-04-2024

NQC System Certificate No. 2414

अंशांकनकर्ता:

Calibrated by:

SURYA

VIKRAM

जांचकर्ता:

Checked by:

Dr. RAJESH KUMAR

जारीकर्ता:

Issued by:

प्रभारी वैज्ञानिक:

Scientist-in-charge:

Dr.S.S.K.TITUS



सी एस आई आर - राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY

(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद्)

(Council of Scientific and Industrial Research)

राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएम एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)

डॉ. के. एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत

Dr. K. S. Krishnan Marg, New Delhi-110012, INDIA

दूरभाष/Phone : 91-11- 4560 8441, 8589, 8610, 9447, फैक्स/ Fax : 91-11- 4560 8448

ई-मेल/ E-mail: cfct@nplindia.org, वेबसाइट/ Website: www.nplindia.org



अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
FORCE PROVING INSTRUMENT

प्रमाण पत्र संख्या/ Certificate No.

2202142 O/D1.05/C-464

दिनांक /Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for the next calibration	पृष्ठ /Page	पृष्ठों की संख्या /No. of Pages
25.02.2022	25.04.2024	2	2

7. Results: Compression

The calibration data obtained is valid for the following digital indicator setting only					
Calibration Signal: Nil			No load output: 32 div.		
(Digital Indicator Reading in Div)					
Applied Force (kN)	Position 0° series 1	Position 0° series 2	Position 120° series 3	Position 240° series 4	Average 1,3,4
0	0	0	0	0	---
150	10038	10033	10032	10029	10033
300	20060	20056	20052	20049	20054
600	40104	40100	40093	40092	40096
900	60132	60126	60123	60121	60125
1200	80149	80141	80145	80144	80146
1500	100160	100154	100168	100156	100161
1800	120168	120159	120173	120165	120169
2100	140165	140155	140171	140167	140168
2400	160162	160148	160175	160158	160165
3000	200119	200115	200147	200120	200129
0	14	11	10	10	--

Interpolation Equation: (Compression)

$$F = -7.6908 \times 10^{-17} .X^3 + 1.6198 \times 10^{-10} .X^2 + 1.4962 \times 10^{-2} .X - 0.1244$$

$$X = 1.5719 \times 10^{-9} .F^3 - 4.8297 \times 10^{-5} .F^2 + 66.8378 .F + 8.3265$$

Where F = Force in kN

X = Indicator reading in div.

Classification: The force proving instrument is found to comply with the requirements of IS: 4169-2014 in respect of interpolated forces and classified as follows:

Class	Mode	From	To	Uncertainty of Measurement
Class 1	Compression	3000 kN	150 kN	± 0.12%

The reported uncertainty is at coverage factor $k=2$ which corresponds to a coverage probability of approximately 95% for a normal distribution, considering the relative deviation of different components such as zero, repeatability, reproducibility, resolution, zero, creep, interpolation and combining with the uncertainty of the applied force.

8. Date of calibration: 25.02.2022

9. Remarks: Nil

अंशांकनकर्ता:

Calibrated by :

SURYA

VIKRAM

जाँचकर्ता:

Checked by :

Dr. RAJESH KUMAR

जारीकर्ता:

Issued by:

प्रभारी वैज्ञानिक:

Scientist-in-charge:

Dr.S.S.K.TITUS



डॉ. प्रीतिवत्स राव शर्मा