



TRUE COPY

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



Visit our Web Site : www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/M/O63/03 Issue No.03

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon - Ambli Road, Ambli, Ahmedabad - 380 058, Gujarat	Certificate No.	NCQC-M/200223/001
	Date of Issue	20/02/2023
	Date of Calibration	20/02/2023
	Suggested Due Date	19/02/2024
	Date of Receipt / Ref.No.:	20/02/2023

ULR No. : CC212823000000755F	Discipline : Mechanical Calibration Mass and Volume
------------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : ===
	Serial No. : ===
	Name of Instrument : Standard Weights

Range	20000 g - 2 Nos	Visual Inspection Accuracy	OK F1 Class
Make / Model Location	Weightronics/=== ===	Std. Ref. Used Cali.Method No.	OIML R 111 - 1 NCQC/CM/M/063
Loc. of Calibration Environment Cond.	At Laboratory 25 ± 1 °C, 40 to 60% Rh, 1000 ± 15 hPa		

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
360	Weights (E2 Class)	Weightronics / =====	NCQC/M - 168	CC-2346	MMD/200822/O4	19/08/2024
Weights (Denomination)		Type	Material	Assume density	Make	
20000 g - 2 Nos.		Cylindrical knob	Stainless steel	7950 ± 140 kg/m ³	Weightronics	

Results of calibration

Sr. No. / Id. No.	Denomination (UUC) in g.	Mass Value in g.	Deviation in g.	Expanded Uncertainty in mg. (k=2)	Status
NCQC/M-98	20000	20000.040	0.040	15.3000	F1
NCQC/M-99	*20000	20000.040	0.040	15.3000	F1

Remarks:

- Mass values of all the weights are conventional mass values and within the maximum errors permissible in "F1" Accuracy Class of weights as per OIML R 111-1.
- Weights are calibrated for scientific or industrial purpose only.
- These results are obtained at the time of calibration.
- Suggested due date is given based on customer requirements.
- Change in temperature and relative humidity during the calibration were less than ± 0.7 °C per hour and ± 10% Rh per 4 hours respectively).
- Calibration is performed on the mass comparator against standards mass by comparison method under controlled conditions (ABBA Method).
- Corrected mass value is calculated based on ABBA method.
- Condition of instrument found satisfactory during receipt.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

Traceable To National / International Standards.

Calibrated By Valid up to 19-02-2024 Reviewed: <i>[Signature]</i>	Himanshu Parbadiya Calibration Engineer	Reviewed & Approved By <i>[Signature]</i>	Jigar Panchal Calibration Incharge
---	--	--	---------------------------------------

NCQC System Certificate No. 273