



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
 E-mail : ncqc@calibrationlaboratory.in , calibrationlab.ncqc@gmail.com
 Visit our Web Site: www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/010/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat	Certificate No.	NCQC-T/200223/012
	Date of Issue	22/02/2023
	Date of Calibration	20/02/2023
	Suggested Due Date	19/02/2024
Date of Receipt / Ref.No.: 16/02/2023		

ULR No. : CC212823000000767F	Discipline : Thermal Calibration Specific Heat and Humidity
------------------------------	--

Details of Observation of Unit Under Calibration	Identification No. : NCQC/T – 153
	Serial No. : 3352
	Name of Instrument : Temperature & Humidity Data Logger

Range	0 to 99 % Rh	Visual Inspection Accuracy	OK ± 5 % Rh
Resolution	0.1 % Rh		
Make / Model	Elitech/RC-4HC		
Location	Laboratory		

Loc. of Calibration	At Laboratory	Std. Ref. Used	NCQC/CMT/010
Environment Cond.	25 ± 4 °C, 30 to 75 % Rh	Cali.Method No.	

Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
335	Digital thermo Hygro Palm	Rotronic / HP32	NCQC/T-161 / 5181935 / 20281776	CC 2733	30056028	05/07/2023
398	4 wire RTD with Indicator	RTD - Tempsens / TPRT-100, Indicator - Tempsens / TEMPMET-09	RTD-NCQC/T-234 / 22000184, Indicator - NCQC/T-178 / 18K587909	CC-2222	EIE/T/221082	25/09/2023

Calibration Results of Humidity @ 25 °C

Set point on Hybrid Temperature and Humidity Calibrator in % Rh	Reading observed by UUC in % Rh	Reading observed by Master instrument in % Rh	Absolute Error in % Rh	Expanded Uncertainty in % Rh (k =2) ±
20.0	20.7	20.15	0.55	3.20
40.0	40.9	40.18	0.72	3.20
50.0	51.1	50.23	0.87	3.20
70.0	71.4	70.35	1.05	3.20
95.0	94.2	95.55	1.35	3.20

NCQC
Valid up to 19-02-2024
Reviewed: *[Signature]*

Traceable To National / International Standards.

Calibrated By	<i>[Signature]</i> Chintan Patel Calibration Engineer	Reviewed & Approved By	<i>[Signature]</i> Jigar Panchal Calibration Incharge
---------------	---	------------------------	---

NCQC System Certificate No 317



Since 1998

NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9828616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

F/CR/T/D 10/03 Issue No. 02

Name of Customer National Centre for Quality Calibration 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat		Certificate No. NCQC-T/200223/012	
		Date of Issue 22/02/2023	
		Date of Calibration 20/02/2023	
		Suggested Due Date 19/02/2024	
		Date of Receipt / Ref.No.: 16/02/2023	

ULR No. : CC212823000000767F **Discipline :** Thermal Calibration
Specific Heat and Humidity

Details of Observation of Unit Under Calibration

Identification No. : NCQC/T – 153
Serial No. : 3352
Name of Instrument : Temperature & Humidity Data Logger

Range	-30 to 60 °C	Accuracy	± 0.5 °C up to 20 to 40°C & Other ±1°C
Least Count	0.1 °C		

Calibration Results of Temperature

Set point on Hybrid Temperature and Humidity Calibrator in °C	Reading observed by UUC in °C	Reading observed by Master instrument in °C	Absolute Error in °C	Expanded Uncertainty in °C (k = 2) ±
-25.0	-25.3	-25.103	0.197	0.78
-15.0	-15.3	-15.117	0.183	0.78
0.0	-0.1	-0.033	0.067	0.78
10.0	10.3	10.187	0.113	0.78
30.0	30.3	30.138	0.162	0.78
50.0	50.4	50.175	0.225	0.78

Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements
- Calibration Points are given based on customer requirements.
- Condition of instrument found satisfactory during receipt.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.
- Portable Hybrid temperature and Humidity calibrator was used as a source of generating temperature and humidity as well as controlling temperature at specified set point.

Traceable To National / International Standards

Calibrated By Chintan Patel Calibration Engineer	Reviewed & Approved By Jigar Panchal Calibration Incharge
---	--