



TRUE COPY

FINIKS ENGINEERING COMPANY

Plot No. 21, Anand Nagar, Near Wasan Automobiles, Sahyadri Chowk, M.I.D.C., Anand Nagar - 414111, [M.S.] Tel. [0241] 2778918, E-mail : finiks97@rediffmail.com



Calibration Certificate

ULR : CC2267 23000007263F

CALIBRATE D ON	ISSUED ON	CERTIFICATE NO.	NEXT CALIBRATION DUE ON (suggested by customer)	PAGE	NO. OF PAGES
13/03/2023	13/03/2023	23-MCC-13-12-002	12/03/2025	1	1

CALIBRATED FOR : M/s National Centre For Quality Calibration,
4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad. 380058

CUSTOMER'S CHALLAN NO. : 060323/009 Date of Receipt : 13/03/2023
Dated : 06/03/2023 Work Order No. : 23-MC-13-12

TYPE OF GAUGE : LENGTH BAR

IDENTIFICATION : MAKE : -- Sr. No. : --
SIZE : 150.0, 200.0 & 300.0 mm ID. No. : NCQC/M-248, 249 & 250

LAB ID NO. : 23-MC-13-12-002 CONDITION AT RECEIPT : OK
ENVIRONMENTAL CONDITION : TEMPERATURE : 20°C (± 1°C) HUMIDITY : 50 ± 10%RH

REFERENCE DOCUMENT USED :

CALIBRATION PROCEDURE : CP/29/00 CALIBRATION LOCATION : In Lab

Equipment / Reference Standards Used For Calibration : Traceable to National / International standards.

Equipment / Reference Standard	Identification No.	Certificate No.	Calib. Date	Valid up to
GRANITE SURFACE PLATE	FINIKS/GR/SP/01	22-MGC-08-01-001	08/07/2022	07/07/2023
ELECTRONIC HEIGHT GAUGE (Digital Linear Master)	FINIKS/HM-01	CGIL/2022/1121/K-003	21/11/2022	20/11/2024
Steel Length Bars	FINIKS/LB/01	LB/21/5849	31/03/2021	31/03/2023

UNCERTAINTY OF MEASUREMENT : ± 5.7 µm

The reported uncertainty is at coverage factor K=2 which corresponds to a coverage probability of approximately 95.45% for normal distribution.

Dimension (Gauges)

CALIBRATION RESULTS

Sr. No.	PARAMETER	NOMINAL VALUE (mm)	OBSERVED VALUE (mm)
1	Length	150.000	150.0032
2	Length	200.000	200.0030
3	Length	300.000	300.0038

x-x-x-x-x-x- End of certificate -x-x-x-x-x-x


NCQC
Valid up to 12-03-2025
Reviewed *[Signature]*

Calibrated by *[Signature]*

Authorized Signatory

[Signature]

This Certificate refers only to the item (s) submitted for calibration. This calibration results in this certificate are valid at the stated conditions of measurements. This certificate shall not be reproduced, except in full, without the permission in writing.



FM/25/00

NCQC System Certificate No. 442